SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

DG-18 Agar

Product Number: 16-DG180 Fill Volume: 60 mL

Lot Number: 2607200

Expiration:10-20-26

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth
E. coli ATCC®25922	Inhibited
B. subtilis ATCC®6633	Inhibited
C. albicans ATCC®10231	Growth
A. brasiliensis ATCC®16404	Growth

Microbial testing was performed aerobically at 35-37°C and ambient temperature for 72 hours
Sterility testing was performed at 35-37° C for a minimum of 24 hours

Growth approved by:_____

(QA Manager)

Microbial Load

Lot Size:	3.5L
Number of Units Sampled:	30
Final pH at 25°C:	5.9±0.2

Appearance

Prepared DG18 Agar is dark to light amber

Lot approved by:_____

(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.