SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Symphony Agar

Product Number: 16-Symph | Expiration: | Fill Volume: 55mL
Lot Number: 2608030 | 9/3/2026

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth	Reaction
C. albicans ATCC 10231	Growth	Round, pink colonies
A. brasiliesis ATCC 16404	Growth	Irregu;ar, green colonies
E. coli ATCC 25922	Inhibited	N/A

Microbial testing was performed aerobically at 35-37°C for 18-24 hours
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by:_____
(QA Manager)

Microbial Load

Lot Size:	2L
Number of Units Sampled:	20
Final pH at 25°C:	5.6±0.2

Appearance

Limpid, violet agar

Lot approved by:_____
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.