SMITH RIVER BIOLOGICA

Quality Biologicals and Prepared Media for the Laboratory Profe

Certificate of Analysis

Anaerobic Blood Agar

Product Number: 11-ABA00 Expiration: 12/31/2026 Fill Volume:
Lot Number: 2608310 Fill Depth:

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth
F. nucrophorum ATCC® 25286	Growth
P. anaerobius ATCC® 27337	Growth
S. aureus ATCC® 25923	Growth at 72h

Microbial testing was performed anaerobically at 35-37°C for 18-24 hours
Sterility check was performed at 35-37°C for 24 hours

Growth approved by:_____
(QA Manager)

Microbial Load

Lot Size:	60
Number of Units Sampled:	35
Final pH at 25°C:	7.3±02

Appearance

Uninoculated media is dark red to red brown and opaque

Lot approved by:_____
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.

LS

ssional

18mL
3-5 mm