SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Sabouraud Dextrose Agar with Chloramphenicol

Product Number: 11-SDAC0 Fill Volume: 20 ± 2 mL

Lot Number: 2609150 Fill Depth: 3-5 mm

Expiration:12-15-26

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth
A. brasiliensis ATCC®16404	Growth after 1-5 days
C. albicans ATCC®10231	Growth after 1-2 days
T. interdigitale ATCC®9533	Growth after 3-7 days
E. coli ATCC®25922	Inhibited at 24 hours

Microbial testing was performed aerobically at 20-25°C

Sterility testing was performed at 35-37°C for 24 hours

Growth approved by:_____

(QA Manager)

Microbial Load

Lot Size:	2
Number of Units Sampled:	30
Final pH at 25°C:	5.6±0.2

Appearance

Prepared SDAC is light to dark amber

Lot approved by:_____

(QA Manager)

This product is not intended for diagnosis or treatment of human disease.