SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Standard Methods Agar

Product Number: 11-SMA00 Fill Volume: 20 mL
Lot Number: 2609070 Fill Depth: 3-5mm
Expiration: 12-7-26

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth
E. coli ATCC®25922	Excellent
S. epidermidis ATCC®12228	Excellent
C. albicans ATCC®10231	Excellent
B. subtilis MCM 3010	Excellent
P. aeruginosa ATCC®27853	Excellent

Microbial testing was performed aerobically at 35-37°C for 18-24 hours
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by:_____
(QA Manager)

Microbial Load

Lot Size:	10
Number of Units Sampled:	30
Final pH at 25°C:	7.0±0.2

Appearance

Uninoculated agar is light to medium amber and clear to hazy.

Lot approved by:_____
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.