SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Tryptic Soy Agar

Product Number: 11-TSA00 Expiration: 1/24/2027 Fill Volume: 20 ± 2 mL
Lot Number: 2609240 Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our Quality Assurance Laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth
C. albicans ATCC® 10231	Growth
A. brasiliensis ATCC® 16404	Inhibited
S epidermidis ATCC® 12228	Growth
B. subtilis ATCC® 6633	Growth

Microbial testing of bacteria was performed aerobically at 35-37°C for 18-24 hours
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by:_____
(QA Manager)

Microbial Load

Lot Size:	150
Number of Units Sampled:	20
Final pH at 25°C:	7.3±0.2

Appearance

Prepared TSA with cycloheximide is light to dark amber and slightly hazy

Lot approved by:_____
(QA Manager)

This product is not intended for the diagnosis or treatment of human disease.