SMITH RIVER BIOLOGICALS

Quality Biologicals and Prepared Media for the Laboratory Professional

Certificate of Analysis

Tryptic Soy Agar with 5% Sheep Blood and Streptomycin

Product Number: 11-TSA5%S200
Lot Number:2607270
Expiration:10-27-26

Fill Volume: 20 mL±2 ml
Fill Depth: 3-5 mm

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our quality control laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

Test Organisms	Growth
S. aureus ATCC®25923	Inhibited
S. agalactiae ATCC®12386	Inhibited
S. pyogenes ATCC®19615	Inhibited
S. pneumoniae ATCC®6305	Inhibited
E. coli ATCC®25922	Inhibited

Microbial testing was performed aerobically at 35-37°C for 18-24 hours
Sterility testing was performed at 35-37°C for 24 hours

Growth approved by:_____
(QA Manager)

Microbial Load

Lot Size:	10L
Number of Units Sampled:	30
Final pH at 25°C:	7.3±0.2

Appearance

Uninoculated plates are bright red to dark red and opaque

Lot approved by:_____
(QA Manager)

This product is not intended for diagnosis or treatment of human disease.